

Marque	Modèle	Temps de traitement du signal (millisecondes)
PHONAK	CLARO 211 AZ	5,3ms
	MAXX 211	5,8ms
	NAÏDA III SP	7,9ms
	AUDEO IX	6,8ms
	AUDEO YES III	7,2ms
UNITRON	NEXT E HP	7,6ms
GN RESOUND	BT4 (analogique)	1,1ms
	CANTA 770D	10,8ms
	Pulse 60	5ms
BELTONE	CHANGE 85D	7,1ms
SIEMENS	8DF	2,9ms
STARKEY	DESTINY 400	4,0ms
	ZÖN 1	3,5ms
BERNAFON	AVANTI 106	6,5ms
WIDEX	LOGO 12 (analogique)	0,6ms
	SENSO C8+	1,5ms
	SENSO DIVA 19	4,1ms
	MIND 440-m	4,3ms
	Passion 440	2,1ms